



Product Service

# EC Certificate

## Product Quality Assurance System

Directive 93/42/EEC on Medical Devices (MDD), Annex VI  
(Devices in class IIa or IIb)

No. G3 15 06 59142 009

**Manufacturer:** EIC Umwelt- und Medizintechnik Ltd.

Rudolf-Diesel-Strasse 5  
37308 Heilbad Heiligenstadt  
GERMANY

**Facility(ies):** EIC Umwelt- und Medizintechnik Ltd.  
Rudolf-Diesel-Strasse 5, 37308 Heilbad Heiligenstadt, GERMANY

**Product Category(ies):** Blood warmer

The Certification Body of TÜV SÜD Product Service GmbH declares that the aforementioned manufacturer has implemented a quality assurance system for final inspection and test of the respective devices / device categories in accordance with MDD Annex VI. This quality assurance system conforms to the requirements of this Directive and is subject to periodical surveillance. For marketing of class IIb devices an additional Annex III certificate is mandatory. See also notes overleaf.

**Report no.:** 713060976

**Valid from:** 2015-07-21

**Valid until:** 2020-05-31

**Date,** 2015-07-22

Hans-Heiner Junker



TÜV SÜD Product Service GmbH is Notified Body with identification no. 0123

Page 1 of 1